




STAR SCIENTIFIC GLASS CO.

(An ISO 9001-2000 Certified)

PIPETTES

S7040 PIPETTES


ARTIFICIAL INSEMINATION, CATTLE, TAPER TIP

Length mm	
450	

These are made from machine-drawn high quality tubing. One end of the pipette is blunt and fire-polished and the other gently tapered and fire polished.


S7056 PIPETTES

BACTERIOLOGICAL, GRADUATED


Capacity ml.	Tolerance \pm ml	
1.1	0.025	
2.2	0.04	

These pipettes are used for the bacteriological examination of milk.

S7057 PIPETTES
GERBER, MILK

Capacity MI.	Tolerance \pm ml	
10.75	0.03	

S7059 PIPETTES
MEASURING (MOHR TYPE) CLASS A


Capacity MI.	Graduation Interval ml	Tolerance \pm ml	
0.1*	0.01	0.006	
0.2*	0.01	0.006	
1	0.01	0.006	
1*	0.1	0.006	
2	0.02	0.01	
2*	0.1	0.01	
5	0.05	0.03	
5*	0.1	0.05	
10	0.1	0.05	
25	0.2	0.1	

* Not covered by IS specification.

These pipettes are graduated for delivery from zero mark to the last graduation mark.

S7060 PIPETTES
 MEASURING (MOHR TYPE) , CLASS B


Capacity MI.	Graduation Interval ml	Tolerance \pm ml
0.1*	0.01	0.01
0.2*	0.01	0.01
1	0.01	0.01
1*	0.1	0.01
2	0.02	0.02
2*	0.1	0.02
5	0.05	0.05
5*	0.1	0.1
10	0.1	0.1
25	0.2	0.2



* Not covered by IS specification.
 These pipettes are graduated for delivery from zero mark to the last graduation mark.


S7062 PIPETTES
 MEASURING (MOHR TYPE) , CLASS B (WHITE MARKING)

Capacity MI.	Graduation Interval ml	Tolerance \pm ml
0.1*	0.01	0.01
0.2*	0.01	0.01
1	0.01	0.01
1*	0.1	0.01
2	0.02	0.02
2*	0.1	0.02
5	0.05	0.05
5*	0.1	0.1
10	0.1	0.1
25	0.2	0.2




* Not covered by IS specification.
 These pipettes are graduated for delivery from zero mark to the last graduation mark.

S7079 PIPETTES
SEROLOGICAL, CLASS A

Capacity MI.	Graduation Interval ml	Tolerance \pm ml	
0.1	0.01	0.006	
0.2	0.01	0.006	
1	0.01	0.006	
1	0.1	0.006	
2	0.02	0.01	
2	0.1	0.01	
5	0.05	0.03	
5	0.1	0.05	
10	0.1	0.05	
25	0.2	0.1	

These pipettes are graduated to tip. Two rings at top indicate that these are calibrated to deliver their total capacity when the last drop is blown out.


S7080 PIPETTES
SEROLOGICAL, CLASS B

Capacity MI.	Graduation Interval ml	Tolerance \pm ml	
0.1	0.01	0.01	
0.2	0.01	0.01	
1	0.01	0.01	
1	0.1	0.01	
2	0.02	0.02	
2	0.1	0.02	
5	0.05	0.05	
5	0.1	0.1	
10	0.1	0.1	
25	0.2	0.2	

These pipettes are graduated to tip. Two rings at top indicate that these are calibrated to deliver their total capacity when the last drop is blown out.

S7082 PIPETTES
 SEROLOGICAL (WHITE MARKING) , CLASS B


Capacity MI.	Graduation Interval ml	Tolerance \pm ml
0.1	0.01	0.01
0.2	0.01	0.01
1	0.01	0.01
1	0.1	0.01
2	0.02	0.02
2	0.1	0.02
5	0.05	0.05
5	0.1	0.1
10	0.1	0.1
25	0.2	0.2



These pipettes are graduated to tip. Two rings at top indicate that these are calibrated to deliver their total capacity when the last drop is blown out.


S7090 PIPETTES
 SEROLOGICAL, KAHN

Capacity MI.	Graduation Interval ml	Tolerance \pm ml
0.1	0.005	0.002
0.2	0.01	0.004
0.5	0.02	0.006




Pipettes are calibrated to deliver from zero mark to the graduation mark.

S7100 PIPETTES
TRANSFER, VOLUMETRIC, CLASS A


Capacity MI.	Tolerance \pm ml	
1*	0.007	
2*	0.01	
5	0.015	
10	0.02	
20	0.03	
25	0.03	
50	0.05	
100	0.08	

* not covered by IS specifications.

S7102 PIPETTES
TRANSFER, VOLUMETRIC, CLASS B

Capacity MI.	Tolerance \pm ml	
1	0.015	
2	0.02	
5	0.03	
10	0.04	
20	0.06	
25	0.06	
50	0.1	
100	0.16	

S7105 PIPETTES
OSTWALD - FOLIN

Capacity MI.	
0.5	
1.0	

These pipettes are generally used for blood serum tests. They are calibrated to deliver by blowing out the last drop.
